





Ultracur3D® Biocompatibility

	ISO 10993-5 - Cytotoxicity	ISO 10993-10 - Skin Irritation	ISO 10993-10 - Skin sensitization
Ultracur3D® CK-Line	1)	0	0
Ultracur3D® EL 150	2)	0	•
Ultracur3D® EL 4000	•	0	0
Ultracur3D® EL 60	•	•	•
Ultracur3D® EPD 2006	•	•	•
Ultracur3D® FL 300	2)	•	•
Ultracur3D® FL 60	•	0	0
Ultracur3D® RG 3280	•	0	0
Ultracur3D® RG 9400 B FR	2)	0	0
Ultracur3D [®] RG 35 ³⁾	•	•	•
Ultracur3D® RG 35 B	•	0	0
Ultracur3D® ST 1400	•	•	•
Ultracur3D® ST 45		•	•
Ultracur3D® ST 45 B	•	0	0
Ultracur3D® ST 45 M		0	0
Ultracur3D® ST 7500 G	•	•	•
Ultracur3D® ST 80	•	•	•

O In Progress

Pass

No Pass

The biocompatibility tests were recorded on test specimen of the above referenced product to show compatibility of the material in general. The biocompatibility tests listed are not part of any continuous production protocol. The test assessments reflect only the test specimen and have to be retested on the final product. It remains the responsibility of the device manufacturers and/or end-users to determine the suitability of all printed parts for their respective application.

For notice:

We give no warranties, expressed or implied, concerning the suitability of mentioned product for use in any medical device and pharmaceutical applications. All information contained in this document is given in good faith and is based on sources believed to be reliable and accurate at the date of publication of this document. It is the responsibility of those to whom we supply our products to ensure that any proprietary rights and existing laws and legislation are observed. The certificate is exclusively for our customers and respective competent authorities. It is not intended for publication either in printed or electronic form (e.g. via Internet) by others. Thus, neither partial nor full publication is allowed without written permission.

O Possible on customer request

¹⁾ Passed, but following specific sample preparation procedure (see Product Information)

Passed, but following specific sample preparation procedure (see Biocompatibility in Extended TDS)

³⁾ In addition, passed ISO 10993-11 - Systemic Toxicity Endotoxins and pyrogens detection